

CERTIFICATE OF COMPLIANCE

Certificate Number 2018-03-29; 2018-07-09 (A1)-E215059
Report Reference E215059-D1001-1/A1/C0-UL-UL
Issue Date 2018-03-29; 2018-07-09 (A1)

Issued to: Eppendorf A G
Applicant Company: Barkhausenweg 1
22339 Hamburg Germany

Listed Company: Same as Applicant

This is to certify that representative samples of Laboratory Freezer
CryoCube F740, CryoCube F740i, CryoCube F740iw, CryoCube F740h, CryoCube F740hi, CryoCube F740hiw

Have been investigated by UL in accordance with the Standard(s) indicated on this Certificate.

Standard(s) for Safety: UL 61010-1, 3rd Edition, May 11, 2012, Revised July 15 2015, CAN/CSA-C22.2 No. 61010-1-12, 3rd Edition, Revision dated July 2015

Additional Standards: UL 471 (edition Tenth 06/17/2016), Sections 42, 44, 49 and 68 and CSA C22.2 No. 120-13 (March 2013) Sections 6.11, 6.3, 6.4 and 6.26

Models F740hi, F740hiw and F740h were investigated to UL 471 Tenth Edition, with revisions through December 8, 2016, and the Canadian Standard for Refrigeration Equipment, CAN/CSA C22.2 No. 120-13 dated March 2013 under report SA45018 Vol1, Sec. 1.


Fluid-containing parts of refrigeration systems meet the relevant pressure-related requirements of IEC 60335-2-89.


Additional Information: See the UL Online Certifications Directory at www.ul.com/database for additional information.

Only those products bearing the UL Certification Mark should be considered as being covered by UL's Certification and Follow-Up Service.

Look for the UL Certification Mark on the product.

This is to certify that representative samples of the product as specified on this certificate were tested according to the current UL requirements.


Bruce Mahrenholz, Assistant Chief Engineer, Global Inspection and Field Services, UL LLC
Joseph Hosey, General Manager, Director of Sales – Canada, UNDERWRITERS LABORATORIES OF CANADA INC.


Helena Y. Wolf, Director, Global Market Access Operations, UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative www.ul.com/contactus





Bruce Mahrenholz

Bruce Mahrenholz, Assistant Chief Engineer, Global Inspection and Field Services, UL LLC
Joseph Hosey, General Manager, Director of Sales – Canada, UNDERWRITERS LABORATORIES OF CANADA INC.

Helena Y. Wolf

Helena Y. Wolf, Director, Global Market Access Operations, UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative www.ul.com/contactus

